

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: APRIL 2017
KIT DESCRIPTION: Maternal Control 2
CATALOGUE NO: MSS5025
LOT NUMBERS: 6471MS
EXPIRY DATE: 2019-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (KU/I)	6471MS	46.9 - 70.3	58.6	61.0	N/A
				59.9	
Free hCG (U/I)	6471MS	30.4 - 50.6	40.5	44.7	N/A
				42.7	

QC TESTED BY: James Breen
DATE: 05-Apr-17
RESULT (PASS/FAIL): PASS
QC APPROVED BY: _____

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) RAe Man. R J Fitzgerald. W Fitzgerald OBE. P F Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016