

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-09-28
KIT DESCRIPTION:	Cystatin C Calibrator Set
CATALOGUE NO:	CYS2699
BATCH NO:	448244
EXPIRY DATE:	2019-05-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II/ 1244CY	0.70	0.60 – 0.81	0.69	mg/L
Level III/ 1245CY	3.30	2.97 – 3.63	3.28	mg/L

APPROVED BY:

Uck

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com