



CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: March 2017
KIT DESCRIPTION: MATERNAL CONTROL LEVEL 1
CATALOGUE NO: MSS5024
LOT NUMBERS: 6450MS
EXPIRY DATE: 2019-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000 XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	6450MS	11.5 - 17.3	14.4	13.1	N/A
				13.3	
Free Beta hCG (U/L)	6450MS	11.8 - 19.6	15.7	15.6	N/A
				16.8	

QC TESTED BY: Davina Graham
DATE: 02-Mar-17
RESULT (PASS/FAIL): Pass
QC APPROVED BY:

Adele Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016