

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: March 2017
KIT DESCRIPTION: Maternal Control Level 3
CATALOGUE NO: MSS5026
LOT NUMBERS: 6453MS
EXPIRY DATE: 2019-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000 XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	6453MS	88.8 - 133	111	113	N/A
				109	
Free Beta hCG (U/l)	6453MS	65.6 - 109	87.4	96.3	N/A
				100	

QC TESTED BY: Davina Graham
DATE: 10 Mar 2017
RESULT (PASS/FAIL): pass
QC APPROVED BY: *Adèle Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com
Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.