



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04
KIT DESCRIPTION:	LIPOPROTEIN (a) CALIBRATOR SERIES
CATALOGUE NO:	LP3404
BATCH NO:	415844
EXPIRY DATE:	2019-02-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 2353CH	LP(a)	12.00	9.60-14.40	13.17	mg/dL
LEVEL II / 2355CH	LP(a)	17.00	13.80-20.80	18.53	mg/dL

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald CBE, R F Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI5738, VAT Number GB 353 0304 00.