

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2017-12-29

KIT DESCRIPTION: IA SPECIALITY 1 LEV 1

CATALOGUE NO: IAS3113

LOT NUMBERS: 1772EC

EXPIRY DATE: 2019-12-28

SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

| ANALYTE | LOT NUMBER | RANGE | TARGET | RESULT | PRECISION (if applicable) |
|--------------------|---------------|-------------|--------|--------|---------------------------|
| | 1772EC | 75 - 125 | 100 | 90.2 | N/A |
| Anti TG (kU/l) | | | | 95.1 | |
| (10,1) | | | | | |
| | 1772EC | 26.1 - 43.5 | 34.8 | 34.1 | N/A |
| Anti TPO (kU/l) | | | | 33.3 | |
| 7 and 11 0 (110/1) | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

| OC TESTED BY: | Nicola McNamara |
|---------------|-----------------|
| | |

DATE: 27-Apr-18

RESULT (PASS/FAIL): Pass

QC APPROVED BY:

adèle Martin.