

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2017-12-29

IA SPECIALITY 2 LEVEL 1 KIT DESCRIPTION:

CATALOGUE NO: IAS3117

LOT NUMBERS: 1775EC **EXPIRY DATE:**

2019-12-28

SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
	1775EC	18.6 - 31.0	24.8	26.4	N/A
CALCITONIN (ng/l)				25.2	
,					
GASTRIN (pmol/l)	1775EC	11.7 - 19.5	15.6	15.5	N/A
				14.6	

QC TESTED BY:		Naomi Clyde			
DATE:	13-Apr-18				
RESULT (PASS/FAIL):		Pass			
QC APPROVED BY:		adèle Martin.			