

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: September 2017
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS LEVEL 1
CATALOGUE NO: IA3109
LOT NUMBERS: 1714EC
EXPIRY DATE: 2020-04-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (U/ml)	1714EC	7.50-11.2	9.37	9.3	N/A
				9.5	
CEA (ng/ml)	1714EC	4.15-6.23	5.19	5.7	
				5.5	
FPSA (ug/L)	1714EC	0.705-1.18	0.940	0.901	
				0.921	

QC TESTED BY: Davina Graham
DATE: 21 Sep 2017
RESULT (PASS/FAIL): Pass
QC APPROVED BY:

Adele Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Man. R J Fitzgerald, VV Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016