

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

**DATE OF MANUFACTURE:** October 2017  
**KIT DESCRIPTION:** IA PREMIUM PLUS 2  
**CATALOGUE NO:** IA3110  
**LOT NUMBERS:** 1764EC  
**EXPIRY DATE:** 2020-10-28  
**SYSTEM OF ANALYSIS:** Siemens Immulite 2000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	1764EC	45.2 - 67.8	56.5	58.5	N/A
				55.8	
TSH (mU/l)	1764EC	1.85 - 2.77	2.31	1.93	N/A
				2.15	
Prolactin (mU/l)	1764EC	661 - 991	826	844	N/A
				871	

**QC TESTED BY:** Naomi Clyde  
**DATE:** 11-Apr-18  
**RESULT (PASS/FAIL):** Pass  
**QC APPROVED BY:** *Adele Martin*

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
**T** +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016