

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2017-10-29

KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS LEVEL 3

CATALOGUE NO: IA3111

LOT NUMBERS: 1766EC

EXPIRY DATE: 2020-10-28

SYSTEM OF ANALYSIS: SIEMENS CENTAUR XP

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Alphafoetoprotein (kU/l)	1766EC	189 - 289	239	239.5	N/A
				244.4	
	1766EC	39.9 - 75.7	57.8	58.1	N/A
CEA (ug/l)				60.2	
total PSA (ug/l)	1766EC	14.5 - 24.1	19.3	20.13	N/A
				19.72	

OC TESTED BY: D	avina	Graham
-----------------	-------	--------

DATE: 23-May-18

RESULT (PASS/FAIL): Pass

QC APPROVED BY:

adèle Martin.