

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2017-10-29
KIT DESCRIPTION: IA PREMIUM PLUS 2
CATALOGUE NO: IA3110
LOT NUMBERS: 1754EC
EXPIRY DATE: 2020-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	1754EC	45.2 - 67.8	56.5	52.8	N/A
				54.7	
C-peptide (nmol/l)	1754EC	1.32 - 2.20	1.76	1.794	N/A
				1.764	
Prolactin (mU/l)	1754EC	661 - 991	826	869	N/A
				869	

QC TESTED BY: Nicola McNamara
DATE: 23-Apr-18
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016