



CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2017-01-28
KIT DESCRIPTION: Tumour Marker Control - Level 2
CATALOGUE NO: TU5002
LOT NUMBERS: 207TU
EXPIRY DATE: 2019-01-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CA 125 (kU/l)	207TU	35.8 - 53.6	44.7	44	N/A
				44	
CA15-3 (kU/l)	207TU	48.0 - 72.0	60	57	N/A
				58	
CEA (ug/l)	207TU	4.46 - 6.68	5.57	5.6	N/A
				5.6	

QC TESTED BY: Davina Graham

DATE: 23 March 2017

RESULT (PASS/FAIL): PASS

QC APPROVED BY: *Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016