



CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2017-01-28
KIT DESCRIPTION: Tumour Marker Control - Level 3
CATALOGUE NO: TU5003
LOT NUMBERS: 208TU
EXPIRY DATE: 2019-01-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CA 125 (kU/l)	208TU	114 - 172	143	133	N/A
				147	
CA15-3 (kU/l)	208TU	127 - 191	159	146	N/A
				156	
CEA (ug/l)	208TU	41.1 - 61.7	51.4	50.6	N/A
				51.3	

QC TESTED BY: Davina Graham
DATE: 23 March 2017
RESULT (PASS/FAIL): PASS
QC APPROVED BY:

Adele Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016