

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2017-10-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS LEVEL 3  
CATALOGUE NO: IA3111  
LOT NUMBERS: 1757EC  
EXPIRY DATE: 2020-10-28  
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
C-PEPTIDE (nmol/l)	1757EC	2.290 - 3.810	3.05	2.979	N/A
				2.972	
PROGESTERONE (nmol/l)	1757EC	54.7 - 82.1	68.4	58.5	N/A
				62.0	
PROLACTIN (mU/l)	1757EC	1283 - 1925	1604	1679	N/A
				1690	

QC TESTED BY: Davina Graham  
DATE: 6-Jul-18  
RESULT (PASS/FAIL): Pass  
QC APPROVED BY: *G. Menter*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S.P. Fitzgerald BSc PhD DSc (Hons) FIAE Man. R.J. Fitzgerald, W. Fitzgerald OBE, R.P. Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016