

## **CERTIFICATE OF ANALYSIS**

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-10-29

KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II - LEVEL 3

CATALOGUE NO: IAS3119

LOT NUMBERS: 1856EC

EXPIRY DATE: 2020-10-28

SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

## **TEST CONTROLS OR VIALS**

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
	1856EC	224 - 374	299	282	N/A
Calcitonin (pg/ml)				286	
**************************************	1856EC	143 - 238	190	191	N/A
Gastrin (pmol/l)				199	

QC TESTED BY:		Davina Graham		
DATE:	12-Dec-19		- JUNEAU AND A	
RESULT (PASS/FAIL):		PASS		
QC APPROVED BY:		a. Martin.		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com