

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-11-29
KIT DESCRIPTION: IA PREMIUM PLUS 2
CATALOGUE NO: IA3110
LOT NUMBERS: 1864EC
EXPIRY DATE: 2021-11-28
SYSTEM OF ANALYSIS: ROCHE COBAS e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CEA (ug/l)	1864EC	16.4 - 24.6	20.5	20.6	N/A
				20.7	
PSA Free (ug/l)	1864EC	5.23 - 8.71	6.97	6.68	N/A
				6.69	
PSA Total (ug/l)	1864EC	9.83 - 16.4	13.1	13.2	N/A
				13.4	

QC TESTED BY: Courtney Smylie
DATE: 20-Sep-19
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Mouton*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016