

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-04-29

KIT DESCRIPTION: Maternal Control

CATALOGUE NO: MSS5024

LOT NUMBERS: 7138MS

EXPIRY DATE: 2021-04-28

SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Alphafoetoprotein kU/l	7138MS	12.1-18.1	15.1	15.6	N/A
				15.7	
Total Beta hCG (mU/ml)	7138MS	4355-6533	5444	5330	N/A
				5295	

QC TEST	ED BY:	Davina Graham	
DATE:	03-Dec-18		
RESULT (PASS/FAIL):		Pass	
QC APPROVED BY:		G-Mentin	

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com