

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-10-29
KIT DESCRIPTION: MATERNAL CONTROL 2
CATALOGUE NO: MSS5025
LOT NUMBERS: 7339MS
EXPIRY DATE: 2021-10-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
PAPP-A (U/l)	7339MS	2.39 - 3.99	3.19	3.23	N/A
				3.18	
Total Beta hCG (U/l)	7339MS	11752 - 17628	14690	14595	N/A
				15244	

QC TESTED BY: Davina Graham
DATE: 6-Jun-19
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number N115738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016