



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| | |
|----------------------|---------------------------|
| DATE OF MANUFACTURE: | 2019-10-29 |
| KIT DESCRIPTION: | Blood Gas Control Level 1 |
| CATALOGUE NO: | BG5001 |
| BATCH NO: | 210BG |
| EXPIRY DATE: | 2020-10-28 |

| SAMPLE | ANALYTE | TARGET | RANGE | RESULT | UNITS |
|------------------|-----------|--------|---------------|--------|--------|
| LEVEL II / 210BG | SODIUM | 121.00 | 115.00-127.00 | 123.27 | mmol/L |
| LEVEL II / 210BG | POTASSIUM | 2.43 | 2.24-2.62 | 2.61 | mmol/L |

APPROVED BY:

Lisa Murray

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE MInst R J Fitzgerald W Fitzgerald OBE R F Kely BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 8304 00.