

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2019-12-29
KIT DESCRIPTION:	CYSTATIN C ELEVATED CONTROL
CATALOGUE NO:	CYS5020
BATCH NO:	1305CY
EXPIRY DATE:	2021-12-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
CYSTATIN C	3.39	3.05 – 3.73	3.35	mg/L

APPROVED BY: S Muelly

QUALITY CONTROL REPRESENTATIVE