

HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2019-01-29	
KIT DESCRIPTION:	TRI-LEVEL CARDIAC CONTROL SET	
CATALOGUE NO:	CQ3259	
BATCH / LOT NO:	514294	
EXPIRY DATE:	2023-01-28	

ANALYTE	TARGET	RANGE	RESULT	UNITS
СКМВ	28.7	23.0-34.4	28.3	U/L
СКМВ	144.0	115.0-173.0	140.5	U/L
CKNAC	86.0	71.0-101.0	80.1	U/L
CKNAC	251.0	206.0-296.0	236.0	U/L
CKNAC	623.0	511.0-735.0	631.3	U/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

APPROVED BY:

Lebecca Jantl

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com