

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE:

2019-10-29

KIT DESCRIPTION:

IMMUNOASSAY SPECIALITY 2 - LEVEL 2

(CHINA)

CATALOGUE NO:

IAS3118

LOT NUMBERS:

1950EC

EXPIRY DATE:

2021-10-28

SYSTEM OF ANALYSIS:

SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

RESULTS FOR TEST CONTROLS OR VIALS					
ANALYTE	LOT NO	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pg/ml)	1950EC	90.4-150	120	124	
				120	
Gastrin (pmol/l)	1950EC	31.1-51.9	41.5	44.1	
				43.1	
PASS or FAIL	PASS				

റ്റ	TEST	חשח	DV.

Aimee Anderson

DATE:

11-Jun-20

RESULT (PASS/FAIL): Pass

QC APPROVED BY:

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com