



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2019-01-29
KIT DESCRIPTION:	TRI-LEVEL CARDIAC CONTROL
CATALOGUE NO:	CQ3259
BATCH NO:	524212
EXPIRY DATE:	2023-01-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 4369CK	CK-NAC	86	71-101	82	U/L
LEVEL II / 4370CK	CK-NAC	251	206-296	239	U/L
LEVEL III / 4371CK	CK-NAC	623	511-735	627	U/L
LEVEL II / 4370CK	CK-MB	28.7	23.0-34.4	29.4	U/L
LEVEL III / 4371CK	CK-MB	144.0	115.0-173.0	146.2	U/L

APPROVED BY: *Louise McArthur*

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MSc R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.