## **RANDOX**

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2019-06-29
KIT DESCRIPTION:	Liquid Protein Calibrator
CATALOGUE NO:	IT2691
BATCH NO:	521162
EXPIRY DATE:	2021-06-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 526LPC	HAPTOGLOBIN	126	101-151	116	mg/dL
LEVEL III/ 510LPC	HAPTOGLOBIN	193	154-232	190	mg/dL

**APPROVED BY:** 

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com