

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-05-29

KIT DESCRIPTION: Immunoassay Premium Plus Level 1

CATALOGUE NO: IA3109

LOT NUMBERS: 1908EC

EXPIRY DATE: 2022-05-28

SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
	1908EC	142 - 236	189	198	N/A
Cortisol (nmol/l)				190	
	1908EC	1.17 - 1.75	1.46	1.42	N/A
DHEAS (umol/l)		· · · · · · · · · · · · · · · · · · ·		1.40	
	1908EC	5.42 - 8.14	6.78	7.20	N/A
FSH (mU/ml)			d value valu	7.41	

QC TESTED BY:		Davina Graham		
DATE:	10-Dec-19		.,,	
RESULT (PASS/FAIL):		PASS		
QC APPROVED BY:		Concertion		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com