

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE:

2019-11-29

KIT DESCRIPTION:

Immunoassay Specialty Control 1 Level 3

CATALOGUE NO:

IAS3115

LOT NUMBERS:

1945EC

EXPIRY DATE:

2021-11-28

SYSTEM OF ANALYSIS:

IS: Roche Cobas 8000 and Siemens Immulite

2000

TEST CONTROLS OR VIALS

RESULTS FOR TEST CONTROLS OR VIALS				
ANALYTE	LOT NO	RANGE	TARGET	RESULT
PTH (pmol/L)	1945EC	119-199	159	140
				145
INSULIN (mU/L)	1945EC	30.0-50.0	40.0	39.2
				38.9
PASS or FAIL	PASS			*

QC TESTED BY:		Emer Ni Bhraonain		
DATE:	09-Jun-20			
RESULT	(PASS/FAIL):	Paşs		
QC APPI	ROVED BY:	Loune relle		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com