

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-11-29
KIT DESCRIPTION:	COAGULATION CONTROL LEVEL 2
CATALOGUE NO:	CG5022
BATCH NO:	303CG
EXPIRY DATE:	2021-11-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 303CG	PT HS+	18.0	N/A	17.3	SECS
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY: LOUIS Alle

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com