

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-05-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 2
CATALOGUE NO: IA3110
LOT NUMBERS: 1917EC
EXPIRY DATE: 2022-05-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Cortisol (nmol/l)	1917EC	412 - 686	549	510	N/A
				499	
DHEA-S ($\mu\text{mol/l}$)	1917EC	8.56 - 12.8	10.7	10.3	N/A
				9.8	
T Uptake (%U)	1917EC	28.5 - 41.1	34.8	33.6	N/A
				33.2	

QC TESTED BY: Emer Ni Bhraonain and Davina Graham
DATE: 18-Dec-19
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016