

HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2019-06-29
KIT DESCRIPTION:	LIPID CONTROL LEVEL 2
CATALOGUE NO:	LE2662
BATCH / LOT NO:	2723CH
EXPIRY DATE:	2022-06-28

ANALYTE	TARGET	RANGE	RESULT	UNITS
CHOLESTEROL	5.33	4.64-6.02	5.33	mmol/L
LIPOPROTEIN (a)	20.60	16.50-24.70	19.71	mg/dL

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | I www.randox.com