

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2019-10-29		
KIT DESCRIPTION:	Assayed Urine Control Level 2		
CATALOGUE NO:	AU2352		
BATCH NO:	1072UC		
EXPIRY DATE:	2023-10-28		

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II 1072UC	GLUCOSE	2.73	2.18-3.28	2.67	mmol/L

APPROVED BY: Lower Mittel

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com