

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2020-07-29 Direct LDL/HDL CHOLESTEROL CALIBRATOR		
KIT DESCRIPTION:			
CATALOGUE NO:	CH2673		
BATCH NO:	2826CH		
EXPIRY DATE:	2023-07-28		

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
D LDL/HDL CAL/ 2826CH	LDL CHOL	3.53	N/A	3.54	mmol/L
D LDL/HDL CAL/ 2826CH	HDL CHOL	2.04	N/A	2.03	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Smackey

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com