

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-04-29  
KIT DESCRIPTION: MATERNAL CONTROL 1  
CATALOGUE NO: MSS5024  
LOT NUMBERS: 7460MS  
EXPIRY DATE: 2022-04-28  
SYSTEM OF ANALYSIS: ROCHE COBAS E411 & SIEMENS IMMULITE 2000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (IU/ml)	7460MS	12.3 – 18.5	15.4	15.4	N/A
				15.6	
Total Beta hCG (U/l)	7460MS	6823 - 10000	8529	9091	N/A
				8597	

QC TESTED BY: Davina Graham  
DATE: 13-Sept-19  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *G. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016