

## **CERTIFICATE OF ANALYSIS**

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2020-06-29

KIT DESCRIPTION: IMMUNOASSAY SPECIALITY I - LEVEL 3

CATALOGUE NO: IAS3115

LOT NUMBERS: 1989EC

EXPIRY DATE: 2022-06-28

SYSTEM OF ANALYSIS: ROCHE COBAS 8000

## **TEST CONTROLS OR VIALS**

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
	1989EC	805-1341	1073	900	N/A
Anti-TG (kU/l)				883	
	1989EC	25.0-41.6	33.3	28.66	N/A
Insulin (mU/I)				28.80	
		The state of the s			
***************************************					
	4.4				

QC TESTED BY:		Aimee Anderson	
DATE:	15-Oct-20		
RESULT (PASS/FAIL):		PASS	, , , , , , , , , , , , , , , , , , , ,
QC APPROVED BY:		B. Marti	

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD D5c (Hons) FIAE Man. R. J Fitzgerald, W Fitzgerald OBE. R. P Kelly BSc FCA. Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI 15738, VAT Number