

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 29-12-2019
KIT DESCRIPTION: Immunoassay Premium Plus – Level 1
CATALOGUE NO: IA3109
LOT NUMBERS: 1937EC
EXPIRY DATE: 28-12-2022
SYSTEM OF ANALYSIS: Roche Cobas 8000.

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Free PSA (ug/l)	1937EC	0.676 - 1.13	0.901	0.93	N/A
				0.92	
Insulin (mU/l)	1937EC	1.62 - 2.70	2.16	2.16	N/A
				2.09	
Total PSA (ug/l)	1937EC	1.46 - 2.43	1.94	1.91	N/A
				1.94	

QC TESTED BY: Joseph Boachie
DATE: 23-OCT-20
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016