

## **CERTIFICATE OF ANALYSIS**

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-11-29

KIT DESCRIPTION: IMMUNOASSAY SPECIALITY 1 LEVEL 1

CATALOGUE NO: IAS3113

LOT NUMBERS: 1943EC

**EXPIRY DATE:** 2021-11-28

SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

## **TEST CONTROLS OR VIALS**

RESULTS FOR TEST CONTROLS OR VIALS					
ANALYTE	LOT NO	RANGE	TARGET	RESULT	PRECISION (if applicable)
IGF 1 (ug/l)	1943EC	59.1-98.5	78.8	78.8	
				81.1	
PTH (pmol/l)	1943EC	4.04-6.73	5.38	4.8	
				4.8	

QC TESTED BY:		Emer Ni Bhraonain		
DATE:	8-May 20			
RESULT	(PASS/FAIL):			
QC APPF	ROVED BY:	Lown Alle		