



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2021-08</b>
<b>KIT DESCRIPTION:</b>	<b>Serology Lyme Disease Positive Control</b>
<b>CATALOGUE NO:</b>	<b>SR10346</b>
<b>BATCH NO:</b>	<b>063SR</b>
<b>EXPIRY DATE:</b>	<b>2023-05-28</b>

<b>SAMPLE</b>	<b>ANALYTE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
<b>063SR</b>	<b>Lyme IgG</b>	<b>+VE</b>	<b>N/A</b>	<b>+VE</b>	<b>IU</b>
<b>063SR</b>	<b>Lyme IgM</b>	<b>+VE</b>	<b>N/A</b>	<b>+VE</b>	<b>IU</b>

APPROVED BY: *Louise McKeown*

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) | [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MInt. R J Fitzgerald, W Fitzgerald OBE, R F Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.