

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2021-07-29
KIT DESCRIPTION:	Liquid CRP Control (CRP Control)
CATALOGUE NO:	CP2480
BATCH NO:	2830CP
EXPIRY DATE:	2023-01-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level II/ 2830CP	CRP	22.5	18.0-27.0	23.8	mg/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE