



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2021-03-01</b>
<b>KIT DESCRIPTION:</b>	<b>Lipid Control (LPD Control 3)</b>
<b>CATALOGUE NO:</b>	<b>LE2670</b>
<b>BATCH NO:</b>	<b>2880CH</b>
<b>EXPIRY DATE:</b>	<b>2024-02-28</b>

<b>SAMPLE</b>	<b>ANALYTE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
Level III/ 2880CH	HDL	1.87	1.59-2.15	1.87	mmol/L
Level III/ 2880CH	LDL	4.94	4.20-5.68	4.91	mmol/L

**APPROVED BY:**

**QUALITY CONTROL REPRESENTATIVE**

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
**T** +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** [marketing@randox.com](mailto:marketing@randox.com) **I** [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Mure, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.