RANDOX

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2021-08-29
KIT DESCRIPTION:	Coagulation Control Level 1
CATALOGUE NO:	CG5021
BATCH NO:	356CG
EXPIRY DATE:	2023-03-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level I / 356CG	APTT	48.4	38.7-58.1	46.1	Seconds

APPROVED BY: JOHN MADON

QUALITY CONTROL REPRESENTATIVE