

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2020-10-29
KIT DESCRIPTION:	ASO CALIBRATOR
CATALOGUE NO:	LO2306
BATCH NO:	957LO
EXPIRY DATE:	2022-04-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 517LPC	ASO	251	201-301	254	IU/mI
LEVEL III / 566LPC	ASO	363	290-436	368	IU/ml

APPROVED BY: Louise M. Mul

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com