

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2021-03-01
KIT DESCRIPTION:	CRP CALIBRATOR SERIES (CRP CAL)
CATALOGUE NO:	CP2499
BATCH NO:	559643
EXPIRY DATE:	2022-08-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level II / 550LPC	frCRP	46.6	37.3 – 55.9	44.9	mg/L
Level III / 566LPC	frCRP	68.9	55.1 – 82.7	69.1	mg/L

APPROVED BY: Madey

QUALITY CONTROL REPRESENTATIVE