

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-12-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS LEVEL 1
CATALOGUE NO: IA3109
LOT NUMBERS: 2105EC
EXPIRY DATE: 2024-12-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FT3 (pmol/l)	2105EC	3.23 - 5.38	4.30	3.92	N/A
				3.92	
FT4 (pmol/l)	2105EC	9.83 - 16.4	13.1	13.0	N/A
				13.1	
TSH (mIU/l)	2105EC	0.074 - 0.112	0.093	0.089	N/A
				0.088	

QC TESTED BY: Jonathan Campbell
DATE: 17-FEB-22
RESULT (PASS/FAIL): PASS
QC APPROVED BY: A. McArthur

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com
Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W J Fitzgerald OBE, R P Kelly BSc FCA,
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738, VAT Number GB 353 0304 00.