

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-03-01

KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II – LEVEL 2

CATALOGUE NO: IAS3118

LOT NUMBERS: 2070EC

EXPIRY DATE: 2023-02-28

SYSTEM OF ANALYSIS: Immulite_2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pmol/l)	2070EC	26.1 - 43.5	34.8	33.9	N/A
				32.2	
Gastrin (pmol/l)	2070EC	33.3 - 55.5	44.4	39.2	N/A
				44.5	
	ļ				

QC TEST Adrian G			
DATE:	20OCT21		
RESULT	(PASS/FAIL):	Pass	
QC APPROVED BY:		a. Martu	

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr. 5 P. Fitzgerald BSr, PhD DSr, (Hons) FIAE Man. R. J. Fitzgerald, W. Fitzgerald OBE, R. P. Kelly BSc FCA, Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5738, VAT Number, GB 353 0304-00.