

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-03-01  
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II – LEVEL 3  
CATALOGUE NO: IAS3119  
LOT NUMBERS: 2071EC  
EXPIRY DATE: 2023-02-28  
SYSTEM OF ANALYSIS: Immulite\_2000

## TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pmol/l)	2071EC	70.5 - 118	94	88.7	N/A
				88.7	
Gastrin (pmol/l)	2071EC	158 - 263	210	221	N/A
				215	

## QC TESTED BY:

Adrian Grecu

DATE: 20OCT21

RESULT (PASS/FAIL): Pass

QC APPROVED BY:

*A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016