

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-05-29
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY I - LEVEL 2
CATALOGUE NO: IAS3114
LOT NUMBERS: 2194EC
EXPIRY DATE: 2024-05-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
C-Peptide (ng/ml)	2194EC	2.38 – 3.96	3.17	3.07	N/A
				3.07	
Parathyroid Hormone (pmol/l)	2194EC	18.4 – 30.6	24.5	23.8	N/A
				23.7	

QC TESTED BY: Davina Graham
DATE: 06-OCT-22
RESULT (PASS/FAIL): PASS
QC APPROVED BY: G. McArthur

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com
Directors: Dr S.P. Fitzgerald BSc PhD DSc (Hons) FRAE Man. R.J. Fitzgerald, W. Fitzgerald OBE, R.P. Kelly BSc FCA,
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI 5733. VAT Number GB 353 0304 00.

RANDOX

Radox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@radox.com **I** www.radox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCG FRCR, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Radox Laboratories Limited is a company registered in Northern Ireland. Company Number: NI 5738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016