

# RANDOX


## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	2022-03-29
<b>KIT DESCRIPTION:</b>	LIPID CONTROL (LPD CONTROL 3)
<b>CATALOGUE NO:</b>	LE2670
<b>BATCH NO:</b>	2971CH
<b>EXPIRY DATE:</b>	2025-03-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level III / 2971CH	CHOL	8.15	7.09 – 9.21	8.35	mmol/L
Level III / 2971CH	DHDL	2.02	1.72 – 2.32	2.08	mmol/L

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE FInst R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0904 00.