

HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| Date of Manufacture: | 2022-10 | | | |
|----------------------|--|--|--|--|
| Product Description: | PRECISION CHEMISTRY CONTROL PREMIUM PLUS LEVEL 2 | | | |
| Catalogue Number: | UN1557 | | | |
| Batch Number: | 1602UN | | | |
| Expiry Date: | 2026-08 | | | |

| Analyte | Lot No. | Target | Range | | Result | Units |
|---------------|---------|--------|-------|-----|--------|-------|
| TOTAL PROTEIN | 1602UN | 57.91 | N/A | N/A | 61.033 | g/l |
| ALBUMIN | 1602UN | 39.94 | N/A | N/A | 42.193 | g/l |

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Approved By: D. Jary

Date: 04 OCF 2022

Quality Control Representative