



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2022-09-29</b>
<b>KIT DESCRIPTION:</b>	<b>Direct HDL/LDL Cholesterol Calibrator</b>
<b>CATALOGUE NO:</b>	<b>CH2673</b>
<b>BATCH NO:</b>	<b>3018CH</b>
<b>EXPIRY DATE:</b>	<b>2025-09-28</b>

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
CALIBRATOR 3018CH	HDL	1.99	N/A	1.99	mmol/L
CALIBRATOR 3018CH	LDL	4.53	N/A	4.51	mmol/L

**APPROVED BY:**  SignNow e-signature ID: 00af9aecef...  
15-Nov-2022 09:09:00 UTC  
Joshua Potts (Signer)

**QUALITY CONTROL REPRESENTATIVE**

**Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom**  
**T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com**

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 000