



HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2022-08-29
KIT DESCRIPTION:	TRI LEVEL CARDIAC CONTROL (CRD CONTROL 1,2,3)
CATALOGUE NO:	CQ3259
BATCH / LOT NO:	616793
EXPIRY DATE:	2026-08-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 4681CK	CK NAC	89	73-105	87	U/L
LEVEL II / 4682CK	CK NAC	224	184-264	226	U/L
LEVEL III / 4683CK	CK NAC	586	481-691	585	U/L
LEVEL II / 4682CK	CK MB	19.9	15.9-23.9	19.2	U/L
LEVEL III / 4683CK	CK MB	94.9	75.9-114	92.5	U/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MSc R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

4705-QC Rev6
27TH September 2013