

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE:			2022-10-2	29		
KIT DESCRIPTION:	1	MMUNOAS	SAY SPE	CIALITY II -	LEVEL 2	2
CATALOGUE NO:	ĪΑ	S3118				
LOT NUMBERS:	2224EC					
EXPIRY DATE:	2024-10-28					
SYSTEM OF ANALY	SIS:	SIEMENS	IMMULITE	=_2000		

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pg/ml)	2224EC	75.0 - 125	100	122	N/A
				121	
Gastrin (pmol/l)	2224EC	30.3 – 50.5	40.4	39.7	N/A
		00.0 00.0	40.4	38.2	
					-

08-Feb-23		
(PASS/FAIL):	PASS	
ROVED BY:	G. Martin.	
	ED BY: raham 08-Feb-23 (PASS/FAIL): ROVED BY:	raham 08-Feb-23 (PASS/FAIL): PASS

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R. J Fitzgerald, W Fitzgerald OBE. R. P. Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5738, VAT Number GB 353 0304 00.